



**Continuous Air Monitoring System (CAMS) Photoionization Detector Concentration Data (15-min Net Avg)**

3/20/2023 Time	Current Action Level ppbv	CAMS 1 VOC ppmv	CAMS 1 VOC ppbv	3/20/2023 Time	Current Action Level ppbv	CAMS 2 VOC ppmv	CAMS 2 VOC ppbv	3/20/2023 Time	Current Action Level ppbv	CAMS 3 VOC ppmv	CAMS 3 VOC ppbv	3/20/2023 Time	Current Action Level ppbv	CAMS 4 VOC ppmv	CAMS 4 VOC ppbv	Notes
0:00	56	0	0	0:00	63	0.0066749	6.674912	0:00	N/A	0	0	0:00	48	0	0	
0:15	56	0	0	0:15	63	0.0063653	6.3653316	0:15	N/A	0	0	0:15	48	0	0	
0:30	56	0	0	0:30	63	0.0062603	6.2602525	0:30	N/A	0	0	0:30	48	0	0	
0:45	56	0	0	0:45	63	0.0063134	6.3133601	0:45	N/A	0	0	0:45	48	0	0	
1:00	56	0	0	1:00	63	0.0064044	6.4044229	1:00	N/A	0	0	1:00	48	0	0	
1:15	56	0	0	1:15	63	0.0063915	6.3914694	1:15	N/A	0	0	1:15	48	0	0	
1:30	56	0	0	1:30	63	0.0061538	6.1538456	1:30	N/A	0	0	1:30	48	0	0	
1:45	56	0	0	1:45	63	0.005998	5.9980086	1:45	N/A	0	0	1:45	48	0	0	
2:00	56	0	0	2:00	63	0.0060164	6.0163929	2:00	N/A	0	0	2:00	48	0	0	
2:15	56	0	0	2:15	63	0.0060845	6.0844771	2:15	N/A	0	0	2:15	48	0	0	
2:30	56	0	0	2:30	63	0.0058455	5.8455468	2:30	N/A	0	0	2:30	48	0	0	
2:45	56	0	0	2:45	63	0.0058716	5.8715817	2:45	N/A	0	0	2:45	48	0	0	
3:00	56	0	0	3:00	63	0.006107	6.1069996	3:00	N/A	0	0	3:00	48	0	0	
3:15	56	0	0	3:15	63	0.0060288	6.028757	3:15	N/A	0	0	3:15	48	0	0	
3:30	56	0	0	3:30	63	0.0060718	6.0717642	3:30	N/A	0	0	3:30	48	0	0	
3:45	56	0	0	3:45	63	0.0060377	6.0376952	3:45	N/A	0	0	3:45	48	0	0	
4:00	56	0	0	4:00	63	0.0060534	6.0534245	4:00	N/A	0	0	4:00	48	0	0	
4:15	56	0	0	4:15	63	0.0061857	6.1856727	4:15	N/A	0	0	4:15	48	0	0	
4:30	56	0	0	4:30	63	0.0059273	5.9273142	4:30	N/A	0	0	4:30	48	1.995E-05	0.0199541	
4:45	56	0	0	4:45	63	0.0056945	5.6944955	4:45	N/A	0	0	4:45	48	8.617E-05	0.0861684	
5:00	56	0	0	5:00	63	0.0057643	5.7643389	5:00	N/A	0	0	5:00	48	0.0003747	0.3746762	
5:15	56	0	0	5:15	63	0.0058545	5.8544781	5:15	N/A	0	0	5:15	48	0.0008283	0.8282701	
5:30	56	0	0	5:30	63	0.0057029	5.7028502	5:30	N/A	0	0	5:30	48	0.0008828	0.8828316	
5:45	56	0	0	5:45	63	0.0052812	5.2811798	5:45	N/A	0	0	5:45	48	0.0008507	0.8506725	
6:00	56	0	0	6:00	63	0.0051991	5.1990941	6:00	N/A	0	0	6:00	48	0.0011484	1.1484275	
6:15	56	0	0	6:15	63	0.0054222	5.4222425	6:15	N/A	0	0	6:15	48	0.001504	1.5039575	
6:30	56	0	0	6:30	63	0.0052762	5.2761946	6:30	N/A	0	0	6:30	48	0.0019882	1.9882216	
6:45	56	0	0	6:45	63	0.0052277	5.227682	6:45	N/A	0	0	6:45	48	0.0027876	2.7875949	
7:00	56	0	0	7:00	63	0.0052617	5.2616891	7:00	N/A	0	0	7:00	48	0.0036073	3.6072667	
7:15	56	0	0	7:15	63	0.0054065	5.4064999	7:15	N/A	0	0	7:15	48	0.0051826	5.1825729	
7:30	56	0	0	7:30	63	0.0054139	5.4139263	7:30	N/A	0	0	7:30	48	0.0066548	6.6547736	
7:45	56	0	0	7:45	63	0.0052719	5.2718768	7:45	N/A	0	0	7:45	48	0.0081073	8.1073429	
8:00	56	0	0	8:00	63	0.0055433	5.5433151	8:00	N/A	0	0	8:00	48	0.0077346	7.7346079	
8:15	56	0	0	8:15	63	0.0055715	5.571524	8:15	N/A	0	0	8:15	48	0.00775	7.7499808	
8:30	56	0	0	8:30	63	0.0049489	4.948914	8:30	N/A	0	0	8:30	48	0.0076542	7.6541714	
8:45	56	0	0	8:45	63	0.0046609	4.6608538	8:45	N/A	0	0	8:45	48	0.0074411	7.4411276	
9:00	56	0	0	9:00	63	0.004679	4.6790422	9:00	N/A	0	0	9:00	48	0.0074	7.4000037	
9:15	56	0.000112	0.1119709	9:15	63	0.0042911	4.2910577	9:15	N/A	0	0	9:15	48	0.0067929	6.7929105	
9:30	56	0.0012077	1.2077005	9:30	63	0.0036364	3.6363745	9:30	N/A	0	0	9:30	48	0.0048591	4.8591111	
9:45	56	0.0031762	3.1762453	9:45	63	0.0036229	3.622888	9:45	N/A	0	0	9:45	48	0.0023206	2.3206348	
10:00	56	0.0050927	5.0927077	10:00	63	0.0041146	4.1146196	10:00	N/A	0	0	10:00	48	0.0013787	1.3786811	
10:15	56	0.006964	6.9640314	10:15	63	0.0046602	4.6602256	10:15	N/A	0	0	10:15	48	0.0009242	0.9242454	
10:30	56	0.0087691	8.7690552	10:30	63	0.0053447	5.34475	10:30	N/A	0	0	10:30	48	0.0012633	1.2632888	

10:45	56	0.0093927	9.3926973	10:45	63	0.0055441	5.544068	10:45	N/A	0	0	10:45	48	0.0018036	1.8036272	
11:00	56	0.0103612	10.361189	11:00	63	0.0058195	5.8194597	11:00	N/A	0.1278519	127.85189	11:00	48	0.0021886	2.1885842	* CAMS 3 Faulty Analyzer Being Replaced
11:15	56	0.0110215	11.021476	11:15	63	0.0063695	6.3695451	11:15	N/A	0.5434577	543.45767	11:15	48	0.0027435	2.7435103	* CAMS 3 Faulty Analyzer Being Replaced
11:30	56	0.0118594	11.859404	11:30	63	0.0069295	6.9294881	11:30	N/A	0	0	11:30	48	0.0035093	3.5093253	
11:45	56	0.0118978	11.897773	11:45	63	0.0069032	6.9032051	11:45	N/A	0.1278575	127.85748	11:45	48	0.0040236	4.0235795	* CAMS 3 Faulty Analyzer Being Replaced
12:00	56	0.0121584	12.158437	12:00	63	0.0074723	7.472324	12:00	N/A	0.5434657	543.46566	12:00	48	0.0042131	4.2131152	* CAMS 3 Faulty Analyzer Being Replaced
12:15	56	0.0119435	11.943546	12:15	63	0.0080328	8.0327755	12:15	N/A	0	0	12:15	48	0.0046603	4.6029892	
12:30	56	0.0117623	11.762347	12:30	63	0.0085102	8.5102153	12:30	N/A	0	0	12:30	48	0.0058288	5.8288384	
12:45	56	0.0115104	11.510391	12:45	63	0.0091746	9.1746227	12:45	N/A	0	0	12:45	48	0.0065464	6.5464498	
13:00	56	0.0112364	11.236363	13:00	63	0.0106294	10.629444	13:00	N/A	0	0	13:00	48	0.0075802	7.5802202	
13:15	56	0.0102431	10.243057	13:15	63	0.0116281	11.628136	13:15	N/A	0	0	13:15	48	0.0084424	8.4424452	
13:30	56	0.0106416	10.64157	13:30	63	0.0126046	12.604568	13:30	N/A	0	0	13:30	48	0.0094495	9.4494832	
13:45	56	0.0106811	10.681145	13:45	63	0.0128667	12.866689	13:45	N/A	0	0	13:45	48	0.0104497	10.449746	
14:00	56	0.0108558	10.85577	14:00	63	0.0138844	13.884382	14:00	N/A	0	0	14:00	48	0.0107309	10.730908	
14:15	56	0.0112536	11.253583	14:15	63	0.0152434	15.243407	14:15	N/A	0	0	14:15	48	0.0113111	11.311056	
14:30	56	0.0109815	10.981534	14:30	63	0.0161453	16.145293	14:30	N/A	0	0	14:30	48	0.0119519	11.951874	
14:45	56	0.0109294	10.929431	14:45	63	0.0170091	17.009071	14:45	N/A	0	0	14:45	48	0.012351	12.351002	
15:00	56	0.0111453	11.145329	15:00	63	0.017519	17.518952	15:00	N/A	0	0	15:00	48	0.0133202	13.320193	
15:15	56	0.0108291	10.829078	15:15	63	0.0174957	17.495682	15:15	N/A	0	0	15:15	48	0.0140024	14.002353	
15:30	56	0.0111588	11.158828	15:30	63	0.0172011	17.201092	15:30	N/A	0	0	15:30	48	0.0142039	14.203909	
15:45	56	0.0113441	11.344141	15:45	63	0.0177467	17.746708	15:45	N/A	0	0	15:45	48	0.0144664	14.466371	
16:00	56	0.0107155	10.715501	16:00	63	0.0175803	17.580326	16:00	N/A	0	0	16:00	48	0.0147903	14.790316	
16:15	56	0.0104347	10.434749	16:15	63	0.0188452	18.845157	16:15	N/A	0	0	16:15	48	0.0145918	14.591771	
16:30	56	0.0094473	9.4472712	16:30	63	0.0196153	19.615293	16:30	N/A	0	0	16:30	48	0.0148598	14.859767	
16:45	56	0.0090945	9.0944595	16:45	63	0.0200676	20.067567	16:45	N/A	0	0	16:45	48	0.0150497	15.049722	
17:00	56	0.0099799	9.9799393	17:00	63	0.0205112	20.511221	17:00	N/A	0	0	17:00	48	0.0154311	15.431131	
17:15	56	0.0088444	8.8443537	17:15	63	0.022066	22.065954	17:15	N/A	0	0	17:15	48	0.0159463	15.946262	
17:30	56	0.0085677	8.5677399	17:30	63	0.0221868	22.186769	17:30	N/A	0	0	17:30	48	0.0162851	16.285075	
17:45	56	0.0081568	8.1568078	17:45	63	0.022611	22.611031	17:45	N/A	0	0	17:45	48	0.0165798	16.579756	
18:00	56	0.0073343	7.3342963	18:00	63	0.0233042	23.304216	18:00	N/A	0	0	18:00	48	0.0171727	17.172731	
18:15	56	0.0065401	6.5401476	18:15	63	0.0240888	24.08878	18:15	N/A	0	0	18:15	48	0.0171037	17.103654	
18:30	56	0.0055267	5.5267483	18:30	63	0.024304	24.303987	18:30	N/A	0	0	18:30	48	0.017304	17.303969	
18:45	56	0.0053128	5.3128005	18:45	63	0.0239679	23.967884	18:45	N/A	0	0	18:45	48	0.0173807	17.380729	
19:00	56	0.0051473	5.1473238	19:00	63	0.0222629	22.262872	19:00	N/A	0	0	19:00	48	0.0162681	16.268053	
19:15	56	0.0050689	5.0688934	19:15	63	0.0197213	19.721346	19:15	N/A	0	0	19:15	48	0.0154651	15.465073	
19:30	56	0.0041763	4.1763382	19:30	63	0.0184856	18.485552	19:30	N/A	0	0	19:30	48	0.0146063	14.606312	
19:45	56	0.0037933	3.7933449	19:45	63	0.0173343	17.334298	19:45	N/A	0	0	19:45	48	0.0135707	13.570726	
20:00	56	0.0032972	3.2972172	20:00	63	0.016111	16.111017	20:00	N/A	0	0	20:00	48	0.0124989	12.49887	
20:15	56	0.002634	2.6340487	20:15	63	0.0150386	15.038556	20:15	N/A	0	0	20:15	48	0.0116102	11.610223	
20:30	56	0.0026532	2.6531744	20:30	63	0.0138662	13.86619	20:30	N/A	0	0	20:30	48	0.0110639	11.063893	
20:45	56	0.0020099	2.0098942	20:45	63	0.0132797	13.279749	20:45	N/A	0	0	20:45	48	0.0108908	10.890765	
21:00	56	0.0019166	1.9166069	21:00	63	0.0123264	12.326383	21:00	N/A	0	0	21:00	48	0.0102541	10.254062	
21:15	56	0.0017277	1.7276959	21:15	63	0.0115219	11.521859	21:15	N/A	0	0	21:15	48	0.0095002	9.5002023	
21:30	56	0.0015043	1.504297	21:30	63	0.0107588	10.758773	21:30	N/A	0	0	21:30	48	0.0091456	9.145553	
21:45	56	0.0011825	1.1825226	21:45	63	0.0101515	10.151476	21:45	N/A	0	0	21:45	48	0.0087838	8.7837852	
22:00	56	0.0008547	0.8547087	22:00	63	0.0099517	9.9516941	22:00	N/A	0	0	22:00	48	0.0085418	8.5417734	
22:15	56	0.0011549	1.1548674	22:15	63	0.009524	9.5240106	22:15	N/A	0	0	22:15	48	0.0083054	8.3054033	
22:30	56	0.0016263	1.6263284	22:30	63	0.0093465	9.346502	22:30	N/A	0	0	22:30	48	0.0080337	8.0336951	
22:45	56	0.0014866	1.4866298	22:45	63	0.0093614	9.3614274	22:45	N/A	0	0	22:45	48	0.0078396	7.8395818	
23:00	56	0.0015669	1.5669431	23:00	63	0.0090553	9.0553121	23:00	N/A	0	0	23:00	48	0.0077753	7.7753031	
23:15	56	0.0018652	1.865157	23:15	63	0.0087892	8.7892103	23:15	N/A	0	0	23:15	48	0.0076791	7.6790969	
23:30	56	0.0022343	2.2343495	23:30	63	0.0091121	9.1121447	23:30	N/A	0	0	23:30	48	0.0078188	7.8188023	
23:45	56	0.0024357	2.4357245	23:45	63	0.0088928	8.8927764	23:45	N/A	0	0	23:45	48	0.0077501	7.7501254	